

Ordering information

Product	Package	Code No.
Direct Bilirubin L-Type R1 Buffer Solution	R1: 4 x 70 mL	411-23695
Direct Bilirubin L-Type R2 Vanadate Solution	R2: 4 x 18 mL	413-23895
Total Bilirubin L-Type R1 Buffer Solution	R1: 4 x 70 mL	417-23295
Total Bilirubin L-Type R2 Vanadate Solution	R2: 4 x 18 mL	419-23495
Bilirubin Calibrator	CAL: 4 x for 3 mL	419-73295

Hitachi 902 / 904 / 911 / 912

Reagent	Total Bilirubin L-Type	Direct Bilirubin L-Type
ANALYZE		
CH TEST/ TYPE	T-BIL / SERUM	D-BIL / SERUM
ASSAY CODE	2POINT END - 10	2POINT END - 10
ASSAY POINT	15 - 31 - 0 - 0	15 - 31 - 0 - 0
DILUTION		
WAVELENGTH(SUB/MAIN)	546 / 450	546 / 450
SAMPLE VOL. (NORMAL)	10.0 - 0 - 0	10.0 - 0 - 0
(DEC.)		
(INC.)		
ABS.LIMIT		
PROZONE LIMIT	32000 / UPPER	32000 / UPPER
REAGENT VOL. R1	280 - 0 - () - 0	280 - 0 - () - 0
R2	0 - 0 - () - 0	0 - 0 - () - 0
R3	70 - 0 - () - 0	70 - 0 - () - 0
R4	0 - 0 - () - 0	0 - 0 - () - 0
CALIBRATION		Blank Calibration
CALIB TYPE	LINEAR-2-2-0	LINEAR-2-2-0
AUTO CALIBRATION	0	0
SD LIMIT	999,9	999,9
DUPLICATE LIMIT	500	500
SENSITIVITY LIMIT	0	0
S1 ABS LIMIT	-32000 / 32000	-32000 / 32000
COMPENSATED LIMIT		
RANGE		
UNIT	mg/dL	mg/dL
INSTRUMENT a	1.0	1.0
FACTOR (Y=aX+b) b	0.0	0.0
STD CONC.	CONC.-POS.-VOL.-PREDIL	CONC.-POS.-VOL.-PREDIL
1	0.0 - 1 - 10 - 0	0.0 - 1 - 10 - 0
2	*1 - *2 - 10 - 0	*1 - *2 - 10 - 0
3	0 - 0 - 0 - 0.0	0 - 0 - 0 - 0.0
4	0 - 0 - 0 - 0.0	0 - 0 - 0 - 0.0
5	0 - 0 - 0 - 0.0	0 - 0 - 0 - 0.0
6	0 - 0 - 0 - 0.0	0 - 0 - 0 - 0.0
CALIBRATION LIST		
K-FACTOR		K-Factor of T-BIL
SYSTEM PARAMETER		
LINEARITY CHECK (4-8)		
(9-)		

*1: Input the bottle size.

0309D1/vk

*2: Input the assigned value of the calibrator.