

β-GLUCAN TEST EARLY DETECTION OF INVASIVE FUNGAL INFECTION

FOR DETERMINING (1→3)-β-D-GLUCAN BY A KINETIC TURBIDIMETRIC ASSAY

- + Single test assay
- + Ready to use reagents
- + User-friendly system



INTRODUCTION

Invasive fungal diseases, are a significant worldwide health problem, and their prevalence is increasing. These opportunistic infections affect immunocompromised patients, those undergoing intensive-care treatment and people with chronic disorders, in particular lung diseases. Invasive fungal diseases are important causes of morbidity and mortality and difficult to diagnose. The early recognition and diagnosis of mycoses is of outstanding importance for improving patient outcomes. However, traditional diagnostic tools such as pathologic histological and fungal cultures lack the sensitivity and capacity needed for early diagnoses.

In most pathogenic fungi, (1→3)-β-D-glucan is an integral component of the cell wall (Fig. 1). Small quantities are released into the blood during infection. The Limulus reagent (LAL: Limulus Amebocyte Lysate), made from the extract of blood cells of

horseshoe crabs, has drawn attention as an *in vitro* diagnostic reagent for mycosis. It reacts with (1→3)-β-D-glucan as well as with endotoxin. The β-Glucan Test exclusively measures the (1→3)-β-D-glucan concentration through a kinetic turbidimetric assay in a sample pretreated with a solution which inactivates endotoxin by the use of a non-ionic detergent and polymyxin B.




Fig. 1: Scheme of fungal cell wall

INTENDED USE

IN VITRO DIAGNOSTIC USE FOR THE QUANTITATIVE DETERMINATION OF (1→3)-β-D-GLUCAN IN SERUM OR PLASMA

KEY FEATURES

- + Single test reagent
- + Calibration by QR code scan
- + Quality control available
- + Simple procedure thanks to ready-to-use reagents and intuitive software
- + Quantitative β-glucan measurement by the kinetic turbidimetric method

TEST PRINCIPLE

Endotoxin in a sample is inactivated by heating the sample at 70°C for 10 minutes with the pretreatment solution, which contains non-ionic detergent and polymyxin B. This pretreatment also deactivates inhibitory protein substances in the sample. When the pretreated sample is mixed with the LAL solution, (1→3)-β-D-glucan in the sample activates Factor G, which initiates the cascade reactions shown in Fig. 2. The turbidity change caused by the gelation reaction is detected as transmittance change. The time taken for the transmittance to reach the threshold value is measured. This interval is defined as gelation time (T_g, Fig. 3).




Fig. 2: Cascade reactions of LAL




Fig. 3: Principle of kinetic turbidimetric method

The log (β-glucan concentration) is in inverse proportion to log [log(T_g)]. When the T_g of an unknown sample is measured, the β-glucan concentration of the sample can be obtained from a standard curve.


TESTING PROCEDURE

PRETREATMENT

Pipet sample
0.1 mL



Incubating
70 °C



Cooling




Pretreatment
reagent

Thermostation

Cooling Station
or ice box

MEASUREMENT

Transfer sample
0.2 mL



Mixing



Pretreated
sample

LAL
reagent

Set the tube in the Toxinometer

INSTRUMENT FEATURES



KINETIC TURBIDIMETRIC ASSAY

- + 16 sample positions, up to 64 with extension modules
- + Measurement starts automatically after sample is inserted
- + Touchscreen display




BARCODE READER

- + Registering reagent and patient information



PRETREATMENT AND LAL REAGENT REQUIRED

- + Easy reagent handling



CALIBRATION WITH BARCODE

- + Reading the calibration data card

- + LIS integration capability
- + Minimal maintenance necessary
- + Dimensions

Toxinometer MT-6500:

350 x 420 x 356 mm (W x D x H), 14 kg

Thermostation TS-70/16:

110 x 260 x 145 mm (W x D x H), 2.5 kg

MT-6500 Extension Module:

190 x 420 x 130 mm (W x D x H), 6.5 kg



TEST FEATURES

CHARACTERISTICS

- + LAL (Limulus Amebocyte Lysate) based test principle
- + Specimen: Serum and plasma
- + Measurement time: Maximum 90 minutes

PERFORMANCE DATA

- + Measurement range: 6 to 600 pg/mL
- + Lentinan used as standard
- + Precision: Maximum CV of 6.6% was observed in within-run experiments
- + Interference: No significant interference observed through bilirubin, hemolysis and antifungal drugs



INSTRUMENTS

CODE	PRODUCT	PACKAGE
993-04701	Toxinometer MT-6500	1 unit
999-04801	MT-6500 Extension Module	1 unit
993-03601	Thermostation TS-70/16	1 unit
998-22211	Cooling Station	1 unit

REAGENTS AND CONSUMABLES

CODE	PRODUCT	PACKAGE
997-04101	β -Glucan Test R2: LAL Reagent	50 x for 0.2 mL
993-04201	β -Glucan Test R1: Pretreatment Solution	50 x 0.9 mL
999-04301	β -Glucan Sample Diluent	10 x 0.9 mL
995-04901	Aluminum Cap	10 x 10 units
995-05001	BC Tip EXT	100 units
991-05101	BC Tip 1000-R	100 units
995-04401	LAL Control	10 x for 0.5 mL

For further information on our products
or to place an order, please contact us.

FUJIFILM Wako Chemicals Europe GmbH
Fuggerstr. 12
41468 Neuss, Germany

Phone +49 2131 311 272
Fax +49 2131 311 110

diagnostics_wkeu@fujifilm.com
www.wako-chemicals.de

