

Ordering information

Reagents	Package	Order No.
LDL-C L-Type R1	R1: 4 x 60 mL	419-24017
LDL-C L-Type R2	R2: 4 x 20 mL	419-24027
LDL-C L-Type R1	R1: 1 x 500 mL	991-24015
LDL-C L-Type R2	R2: 1 x 500 mL	991-24025
LDL-C Calibrator	CAL: 4 x for 1 mL	419-24032

Measuring Temperature: 37°C

Hitachi 902/ 904 / 911/ 912

REAGENT	LDL-C L-Type
ANALYZE	
TEST	LDL-C
ASSAY CODE/ TIME	2 POINT END - 10
ASSAY POINTS	15 - 31 - 0 - 0
WAVELENGTH (SUB / MAIN)	700 / 600
DILUTION	/ 0
SAMPLE VOLUME (NORM.)	3 - 0 - 0
(DEC.)	3 - 0 - 0
(INC.)	3 - 0 - 0
ABS. LIMIT	32000 / 32000 / INCREASE
PROZONE LIMIT	-32000 / 32000 / LOWER
REAGENT R1	270 - 0 - () - 0
R2	0 - 0 - () - 0
R3	90 - 0 - () - 0
R4	0 - 0 - () - 0
CALIBRATION	
CALIBRATION TYPE	LINEAR
NUMBER OF POINTS	2
SPAN POINT	2
AUTO CALIBRATION	
SD LIMIT	999.9
DUPLICATE LIMIT	500
SENSITIVITY LIMIT	-99999 - 999999
S1 ABS LIMIT	-32000 - 32000
COMPENSATED LIMIT	
RANGE	
UNIT	mg/dL
INSTRUMENT FACTOR	(Y=aX+b)
a	1.0
b	0.0
STANDARDS	Cal.Code/Conc./Pos./S.Vol./S.Dil./Dil.
1	Water / 0.0 / - / 3 / 0 / 0
2	(*) / *1 / *2 / 3 / 0 / 0
3	
4	
5	
6	

* : User-defined

*1: Assigned value of the calibrator

*2: Position of the calibrator